


SOLUTION CHANGE

DataMaster DMT: 100154
Location: WASHINGTON COUNTY SO
Date: 08/25/2016
Performed by: MARC P POULIN



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 16-04-100
Range = 0.093 - 0.103
Average = 0.099 g/210L
Std Dev = 0.0005

KD
9/16/16

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "M. Poulin", is written over a horizontal line.

Date

08/25/2016